



2655 Park Center Dr., Suite A
Simi Valley, CA 93065
T: +1 805 526 7161
F: +1 805 526 7270
www.alsglobal.com

LABORATORY REPORT

September 28, 2016

Stephen Dutz
South Coast Air Quality Management District
21865 Copley Drive
Diamond Bar, CA 91765

RE: Exxon Mobile Torrance Refinery ER

Dear Stephen:

Enclosed are the results of the sample submitted to our laboratory on September 27, 2016. For your reference, this analysis has been assigned our service request number P1604560.

All analyses were performed according to our laboratory's NELAP and DoD-ELAP-approved quality assurance program. The test results meet requirements of the current NELAP and DoD-ELAP standards, where applicable, and except as noted in the laboratory case narrative provided. For a specific list of NELAP and DoD-ELAP-accredited analytes, refer to the certifications section at www.alsglobal.com. Results are intended to be considered in their entirety and apply only to the samples analyzed and reported herein.

If you have any questions, please call me at (805) 526-7161.

Respectfully submitted,

ALS | Environmental

By Samantha Henningsen at 10:06 am, Sep 28, 2016

Samantha Henningsen
Project Manager



2655 Park Center Dr., Suite A
Simi Valley, CA 93065
T: +1 805 526 7161
F: +1 805 526 7270
www.alsglobal.com

Client: South Coast Air Quality Management District
Project: Exxon Mobile Torrance Refinery ER

Service Request No: P1604560

CASE NARRATIVE

The sample was received intact under chain of custody on September 27, 2016 and was stored in accordance with the analytical method requirements. Please refer to the sample acceptance check form for additional information. The results reported herein are applicable only to the condition of the sample at the time of sample receipt.

Volatile Organic Compound Analysis

The sample was analyzed for volatile organic compounds and tentatively identified compounds in accordance with EPA Method TO-15 from the Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, Second Edition (EPA/625/R-96/010b), January, 1999. This procedure is described in laboratory SOP VOA-TO15. The analytical system was comprised of a gas chromatograph / mass spectrometer (GC/MS) interfaced to a whole-air preconcentrator. This method is included on the laboratory's NELAP and DoD-ELAP scope of accreditation, however it is not part of the AIHA-LAP, LLC accreditation. Any analytes flagged with an X are not included on the NELAP or DoD-ELAP accreditation.

The results of analyses are given in the attached laboratory report. All results are intended to be considered in their entirety, and ALS Environmental (ALS) is not responsible for utilization of less than the complete report.

Use of ALS Environmental (ALS)'s Name. Client shall not use ALS's name or trademark in any marketing or reporting materials, press releases or in any other manner ("Materials") whatsoever and shall not attribute to ALS any test result, tolerance or specification derived from ALS's data ("Attribution") without ALS's prior written consent, which may be withheld by ALS for any reason in its sole discretion. To request ALS's consent, Client shall provide copies of the proposed Materials or Attribution and describe in writing Client's proposed use of such Materials or Attribution. If ALS has not provided written approval of the Materials or Attribution within ten (10) days of receipt from Client, Client's request to use ALS's name or trademark in any Materials or Attribution shall be deemed denied. ALS may, in its discretion, reasonably charge Client for its time in reviewing Materials or Attribution requests. Client acknowledges and agrees that the unauthorized use of ALS's name or trademark may cause ALS to incur irreparable harm for which the recovery of money damages will be inadequate. Accordingly, Client acknowledges and agrees that a violation shall justify preliminary injunctive relief. For questions contact the laboratory.



2655 Park Center Dr., Suite A
 Simi Valley, CA 93065
 T: +1 805 526 7161
 F: +1 805 526 7270
www.alsglobal.com

ALS Environmental – Simi Valley

CERTIFICATIONS, ACCREDITATIONS, AND REGISTRATIONS

Agency	Web Site	Number
AIHA-LAP, LLC	http://www.aihaaccreditedlabs.org	101661
Arizona DHS	http://www.azdhs.gov/lab/license/env.htm	AZ0694
PJLA (DoD ELAP)	http://www.pjlab.com/search-accredited-labs	65818 (Testing)
Florida DOH (NELAP)	http://www.doh.state.fl.us/lab/EnvLabCert/WaterCert.htm	E871020
Maine DHHS	http://www.maine.gov/dhhs/mecdc/environmental-health/water/dwp-services/labcert/labcert.htm	2014025
Minnesota DOH (NELAP)	http://www.health.state.mn.us/accreditation	977273
New Jersey DEP (NELAP)	http://www.nj.gov/dep/oqa/	CA009
New York DOH (NELAP)	http://www.wadsworth.org/labcert/elap/elap.html	11221
Oregon PHD (NELAP)	http://public.health.oregon.gov/LaboratoryServices/EnvironmentalLaboratoryAccreditation/Pages/index.aspx	4068-003
Pennsylvania DEP	http://www.depweb.state.pa.us/labs	68-03307 (Registration)
Texas CEQ (NELAP)	http://www.tceq.texas.gov/field/qa/env_lab_accreditation.html	T104704413- 16-7
Utah DOH (NELAP)	http://www.health.utah.gov/lab/labimp/certification/index.html	CA01627201 6-6
Washington DOE	http://www.ecy.wa.gov/programs/eap/labs/lab-accreditation.html	C946

Analyses were performed according to our laboratory's NELAP and DoD-ELAP approved quality assurance program. A complete listing of specific NELAP and DoD-ELAP certified analytes can be found in the certifications section at www.alsglobal.com, or at the accreditation body's website.

Each of the certifications listed above have an explicit Scope of Accreditation that applies to specific matrices/methods/analytes; therefore, please contact the laboratory for information corresponding to a particular certification.

ALS ENVIRONMENTAL

DETAIL SUMMARY REPORT

Client: South Coast Air Quality Management District
Project ID: Exxon Mobile Torrance Refinery ER

Service Request: P1604560

Date Received: 9/27/2016
Time Received: 09:50

TO-15 - VOC Cans

Client Sample ID	Lab Code	Matrix	Date Collected	Time Collected	Pi1 (psig)	Pf1 (psig)	
Ermanita/ 98th Ave	P1604560-001	Air	9/23/2016	10:54	-0.29	3.82	X

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT
 INVOICE
 LAP AUD.
LABORATC

WO #: 1627009



P1694560

TO: SCAQMD LAB: OTHER:
SOURCE NAME: Exxon Mobile Torrance Refinery ER I.D. No. _____
Source Address: _____ City: Torrance
Mailing Address: _____ City: _____ Zip: _____
Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 9-23-16
Approved by: Jason Low Office: _____ Budget #: 44716
REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
Suspected Violation Rule(s) _____ Other near source monitoring

Sample Collected by: E Holden / S Boddeker Date: 9-23-16 Time: 10:54-12:55

REQUESTED ANALYSIS: TO-15

Location	Can#	Start day / time/ duration	Start vac	End vac
<i>-01</i> Ermanita / 98 th Ave 33.85981, -118.3323 <i>Background</i>	54037	9-23-16 / 10:54 / ~0.5 min	-30	0 <i>15</i>
<i>-02</i> Crenshaw West of Flare 33.85096, -118.328656 <i>Downwind</i>	E3686	9-23-16 / 12:55 / ~0.5 min	-30	-3

*Sent to ALS via QSO for analysis
- Dutz 09/26/16*

Relinquished by	Received by	Firm/Agency	Date	Time
S Boddeker <i>SB</i>	<i>[Signature]</i>	SCAQMD	9-23-16	19:24
<i>[Signature]</i>	<i>[Signature]</i>	SCAQMD	9-23-16	19:26
<i>[Signature]</i>	<i>[Signature]</i>	SCAQMD	09-26-16	14:21

9/27/16 0950

Remarks: Emergency Flare release along Del Amo

**ALS Environmental
Sample Acceptance Check Form**

Client: South Coast Air Quality Management District Work order: P1604560
 Project: Exxon Mobile Torrance Refinery ER
 Sample(s) received on: 9/27/16 Date opened: 9/27/16 by: KKELPE

Note: This form is used for all samples received by ALS. The use of this form for custody seals is strictly meant to indicate presence/absence and not as an indication of compliance or nonconformity. Thermal preservation and pH will only be evaluated either at the request of the client and/or as required by the method/SOP.

- | | Yes | No | N/A |
|---|-------------------------------------|-------------------------------------|-------------------------------------|
| 1 Were sample containers properly marked with client sample ID? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 2 Did sample containers arrive in good condition? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 3 Were chain-of-custody papers used and filled out? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 4 Did sample container labels and/or tags agree with custody papers? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 5 Was sample volume received adequate for analysis? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 6 Are samples within specified holding times? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
| 7 Was proper temperature (thermal preservation) of cooler at receipt adhered to? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 8 Were custody seals on outside of cooler/Box/Container? | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Location of seal(s)? _____ Sealing Lid? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Were signature and date included? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Were seals intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 9 Do containers have appropriate preservation , according to method/SOP or Client specified information? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Is there a client indication that the submitted samples are pH preserved? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Were VOA vials checked for presence/absence of air bubbles? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Does the client/method/SOP require that the analyst check the sample pH and <u>if necessary</u> alter it? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 10 Tubes: Are the tubes capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| 11 Badges: Are the badges properly capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |
| Are dual bed badges separated and individually capped and intact? | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> |

Lab Sample ID	Container Description	Required pH *	Received pH	Adjusted pH	VOA Headspace (Presence/Absence)	Receipt / Preservation Comments
P1604560-001.01	Client Canister					

Explain any discrepancies: (include lab sample ID numbers): _____

RSK - MEEPP, HCL (pH<2); RSK - CO2, (pH 5-8); Sulfur (pH>4)

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 4

Client: South Coast Air Quality Management District
Client Sample ID: Ermanita/ 98th Ave
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P1604560-001

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: 9/23/16
 Date Received: 9/27/16
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 1.00 Liter(s)

Initial Pressure (psig): -0.29 Final Pressure (psig): 3.82

Canister Dilution Factor: 1.29

CAS #	Compound	Result	MRL	Result	MRL	Data Qualifier
		µg/m ³	µg/m ³	ppbV	ppbV	
115-07-1	Propene	0.65	0.65	0.37	0.37	
75-71-8	Dichlorodifluoromethane (CFC 12)	2.1	0.65	0.43	0.13	
74-87-3	Chloromethane	ND	0.65	ND	0.31	
76-14-2	1,2-Dichloro-1,1,2,2-tetrafluoroethane (CFC 114)	ND	0.65	ND	0.092	
75-01-4	Vinyl Chloride	ND	0.65	ND	0.25	
106-99-0	1,3-Butadiene	ND	0.65	ND	0.29	
74-83-9	Bromomethane	ND	0.65	ND	0.17	
75-00-3	Chloroethane	ND	0.65	ND	0.24	
64-17-5	Ethanol	ND	6.5	ND	3.4	
75-05-8	Acetonitrile	ND	0.65	ND	0.38	
107-02-8	Acrolein	ND	2.6	ND	1.1	
67-64-1	Acetone	10	6.5	4.2	2.7	
75-69-4	Trichlorofluoromethane	1.2	0.65	0.22	0.11	
67-63-0	2-Propanol (Isopropyl Alcohol)	ND	6.5	ND	2.6	
107-13-1	Acrylonitrile	ND	0.65	ND	0.30	
75-35-4	1,1-Dichloroethene	ND	0.65	ND	0.16	
75-09-2	Methylene Chloride	0.66	0.65	0.19	0.19	
107-05-1	3-Chloro-1-propene (Allyl Chloride)	ND	0.65	ND	0.21	
76-13-1	Trichlorotrifluoroethane	ND	0.65	ND	0.084	
75-15-0	Carbon Disulfide	ND	6.5	ND	2.1	
156-60-5	trans-1,2-Dichloroethene	ND	0.65	ND	0.16	
75-34-3	1,1-Dichloroethane	ND	0.65	ND	0.16	
1634-04-4	Methyl tert-Butyl Ether	ND	0.65	ND	0.18	
108-05-4	Vinyl Acetate	ND	6.5	ND	1.8	
78-93-3	2-Butanone (MEK)	ND	6.5	ND	2.2	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 2 of 4

Client: South Coast Air Quality Management District
Client Sample ID: Ermanita/ 98th Ave
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P1604560-001

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: 9/23/16
 Date Received: 9/27/16
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 1.00 Liter(s)

Initial Pressure (psig): -0.29 Final Pressure (psig): 3.82

Canister Dilution Factor: 1.29

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
156-59-2	cis-1,2-Dichloroethene	ND	0.65	ND	0.16	
141-78-6	Ethyl Acetate	ND	1.3	ND	0.36	
110-54-3	n-Hexane	ND	0.65	ND	0.18	
67-66-3	Chloroform	ND	0.65	ND	0.13	
109-99-9	Tetrahydrofuran (THF)	ND	0.65	ND	0.22	
107-06-2	1,2-Dichloroethane	ND	0.65	ND	0.16	
71-55-6	1,1,1-Trichloroethane	ND	0.65	ND	0.12	
71-43-2	Benzene	ND	0.65	ND	0.20	
56-23-5	Carbon Tetrachloride	ND	0.65	ND	0.10	
110-82-7	Cyclohexane	ND	1.3	ND	0.37	
78-87-5	1,2-Dichloropropane	ND	0.65	ND	0.14	
75-27-4	Bromodichloromethane	ND	0.65	ND	0.096	
79-01-6	Trichloroethene	ND	0.65	ND	0.12	
123-91-1	1,4-Dioxane	ND	0.65	ND	0.18	
80-62-6	Methyl Methacrylate	ND	1.3	ND	0.32	
142-82-5	n-Heptane	ND	0.65	ND	0.16	
10061-01-5	cis-1,3-Dichloropropene	ND	0.65	ND	0.14	
108-10-1	4-Methyl-2-pentanone	ND	0.65	ND	0.16	
10061-02-6	trans-1,3-Dichloropropene	ND	0.65	ND	0.14	
79-00-5	1,1,2-Trichloroethane	ND	0.65	ND	0.12	
108-88-3	Toluene	1.3	0.65	0.35	0.17	
591-78-6	2-Hexanone	ND	0.65	ND	0.16	
124-48-1	Dibromochloromethane	ND	0.65	ND	0.076	
106-93-4	1,2-Dibromoethane	ND	0.65	ND	0.084	
123-86-4	n-Butyl Acetate	ND	0.65	ND	0.14	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 3 of 4

Client: South Coast Air Quality Management District
Client Sample ID: Ermanita/ 98th Ave
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P1604560-001

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: 9/23/16
 Date Received: 9/27/16
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 1.00 Liter(s)

Initial Pressure (psig): -0.29 Final Pressure (psig): 3.82

Canister Dilution Factor: 1.29

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
111-65-9	n-Octane	ND	0.65	ND	0.14	
127-18-4	Tetrachloroethene	ND	0.65	ND	0.095	
108-90-7	Chlorobenzene	ND	0.65	ND	0.14	
100-41-4	Ethylbenzene	ND	0.65	ND	0.15	
179601-23-1	m,p-Xylenes	ND	1.3	ND	0.30	
75-25-2	Bromoform	ND	0.65	ND	0.062	
100-42-5	Styrene	ND	0.65	ND	0.15	
95-47-6	o-Xylene	ND	0.65	ND	0.15	
111-84-2	n-Nonane	ND	0.65	ND	0.12	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.65	ND	0.094	
98-82-8	Cumene	ND	0.65	ND	0.13	
80-56-8	alpha-Pinene	ND	0.65	ND	0.12	
103-65-1	n-Propylbenzene	ND	0.65	ND	0.13	
622-96-8	4-Ethyltoluene	ND	0.65	ND	0.13	
108-67-8	1,3,5-Trimethylbenzene	ND	0.65	ND	0.13	
95-63-6	1,2,4-Trimethylbenzene	ND	0.65	ND	0.13	
100-44-7	Benzyl Chloride	ND	0.65	ND	0.12	
541-73-1	1,3-Dichlorobenzene	ND	0.65	ND	0.11	
106-46-7	1,4-Dichlorobenzene	ND	0.65	ND	0.11	
95-50-1	1,2-Dichlorobenzene	ND	0.65	ND	0.11	
5989-27-5	d-Limonene	ND	0.65	ND	0.12	
96-12-8	1,2-Dibromo-3-chloropropane	ND	0.65	ND	0.067	
120-82-1	1,2,4-Trichlorobenzene	ND	0.65	ND	0.087	
91-20-3	Naphthalene	ND	0.65	ND	0.12	
87-68-3	Hexachlorobutadiene	ND	0.65	ND	0.060	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 4 of 4

Client: South Coast Air Quality Management District
Client Sample ID: Ermanita/ 98th Ave
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
ALS Sample ID: P1604560-001

Tentatively Identified Compounds

Test Code: EPA TO-15
Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
Analyst: Evelyn Alvarez
Sample Type: Canister
Test Notes:

Date Collected: 9/23/16
Date Received: 9/27/16
Date Analyzed: 9/27/16
Volume(s) Analyzed: 1.00 Liter(s)

Initial Pressure (psig): -0.29 Final Pressure (psig): 3.82

Canister Dilution Factor: 1.29

GC/MS Retention Time	Compound Identification	Concentration µg/m ³	Data Qualifier
No Compounds Detected			

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 1 of 4

Client: South Coast Air Quality Management District
Client Sample ID: Method Blank
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P160927-MB

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: NA
 Date Received: NA
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 1.00 Liter(s)

Canister Dilution Factor: 1.00

CAS #	Compound	Result	MRL	Result	MRL	Data Qualifier
		µg/m ³	µg/m ³	ppbV	ppbV	
115-07-1	Propene	ND	0.50	ND	0.29	
75-71-8	Dichlorodifluoromethane (CFC 12)	ND	0.50	ND	0.10	
74-87-3	Chloromethane	ND	0.50	ND	0.24	
76-14-2	1,2-Dichloro-1,1,2,2-tetrafluoroethane (CFC 114)	ND	0.50	ND	0.072	
75-01-4	Vinyl Chloride	ND	0.50	ND	0.20	
106-99-0	1,3-Butadiene	ND	0.50	ND	0.23	
74-83-9	Bromomethane	ND	0.50	ND	0.13	
75-00-3	Chloroethane	ND	0.50	ND	0.19	
64-17-5	Ethanol	ND	5.0	ND	2.7	
75-05-8	Acetonitrile	ND	0.50	ND	0.30	
107-02-8	Acrolein	ND	2.0	ND	0.87	
67-64-1	Acetone	ND	5.0	ND	2.1	
75-69-4	Trichlorofluoromethane	ND	0.50	ND	0.089	
67-63-0	2-Propanol (Isopropyl Alcohol)	ND	5.0	ND	2.0	
107-13-1	Acrylonitrile	ND	0.50	ND	0.23	
75-35-4	1,1-Dichloroethene	ND	0.50	ND	0.13	
75-09-2	Methylene Chloride	ND	0.50	ND	0.14	
107-05-1	3-Chloro-1-propene (Allyl Chloride)	ND	0.50	ND	0.16	
76-13-1	Trichlorotrifluoroethane	ND	0.50	ND	0.065	
75-15-0	Carbon Disulfide	ND	5.0	ND	1.6	
156-60-5	trans-1,2-Dichloroethene	ND	0.50	ND	0.13	
75-34-3	1,1-Dichloroethane	ND	0.50	ND	0.12	
1634-04-4	Methyl tert-Butyl Ether	ND	0.50	ND	0.14	
108-05-4	Vinyl Acetate	ND	5.0	ND	1.4	
78-93-3	2-Butanone (MEK)	ND	5.0	ND	1.7	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 2 of 4

Client: South Coast Air Quality Management District
Client Sample ID: Method Blank
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P160927-MB

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: NA
 Date Received: NA
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 1.00 Liter(s)

Canister Dilution Factor: 1.00

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
156-59-2	cis-1,2-Dichloroethene	ND	0.50	ND	0.13	
141-78-6	Ethyl Acetate	ND	1.0	ND	0.28	
110-54-3	n-Hexane	ND	0.50	ND	0.14	
67-66-3	Chloroform	ND	0.50	ND	0.10	
109-99-9	Tetrahydrofuran (THF)	ND	0.50	ND	0.17	
107-06-2	1,2-Dichloroethane	ND	0.50	ND	0.12	
71-55-6	1,1,1-Trichloroethane	ND	0.50	ND	0.092	
71-43-2	Benzene	ND	0.50	ND	0.16	
56-23-5	Carbon Tetrachloride	ND	0.50	ND	0.080	
110-82-7	Cyclohexane	ND	1.0	ND	0.29	
78-87-5	1,2-Dichloropropane	ND	0.50	ND	0.11	
75-27-4	Bromodichloromethane	ND	0.50	ND	0.075	
79-01-6	Trichloroethene	ND	0.50	ND	0.093	
123-91-1	1,4-Dioxane	ND	0.50	ND	0.14	
80-62-6	Methyl Methacrylate	ND	1.0	ND	0.24	
142-82-5	n-Heptane	ND	0.50	ND	0.12	
10061-01-5	cis-1,3-Dichloropropene	ND	0.50	ND	0.11	
108-10-1	4-Methyl-2-pentanone	ND	0.50	ND	0.12	
10061-02-6	trans-1,3-Dichloropropene	ND	0.50	ND	0.11	
79-00-5	1,1,2-Trichloroethane	ND	0.50	ND	0.092	
108-88-3	Toluene	ND	0.50	ND	0.13	
591-78-6	2-Hexanone	ND	0.50	ND	0.12	
124-48-1	Dibromochloromethane	ND	0.50	ND	0.059	
106-93-4	1,2-Dibromoethane	ND	0.50	ND	0.065	
123-86-4	n-Butyl Acetate	ND	0.50	ND	0.11	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 3 of 4

Client: South Coast Air Quality Management District
Client Sample ID: Method Blank
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P160927-MB

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: NA
 Date Received: NA
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 1.00 Liter(s)

Canister Dilution Factor: 1.00

CAS #	Compound	Result µg/m ³	MRL µg/m ³	Result ppbV	MRL ppbV	Data Qualifier
111-65-9	n-Octane	ND	0.50	ND	0.11	
127-18-4	Tetrachloroethene	ND	0.50	ND	0.074	
108-90-7	Chlorobenzene	ND	0.50	ND	0.11	
100-41-4	Ethylbenzene	ND	0.50	ND	0.12	
179601-23-1	m,p-Xylenes	ND	1.0	ND	0.23	
75-25-2	Bromoform	ND	0.50	ND	0.048	
100-42-5	Styrene	ND	0.50	ND	0.12	
95-47-6	o-Xylene	ND	0.50	ND	0.12	
111-84-2	n-Nonane	ND	0.50	ND	0.095	
79-34-5	1,1,2,2-Tetrachloroethane	ND	0.50	ND	0.073	
98-82-8	Cumene	ND	0.50	ND	0.10	
80-56-8	alpha-Pinene	ND	0.50	ND	0.090	
103-65-1	n-Propylbenzene	ND	0.50	ND	0.10	
622-96-8	4-Ethyltoluene	ND	0.50	ND	0.10	
108-67-8	1,3,5-Trimethylbenzene	ND	0.50	ND	0.10	
95-63-6	1,2,4-Trimethylbenzene	ND	0.50	ND	0.10	
100-44-7	Benzyl Chloride	ND	0.50	ND	0.097	
541-73-1	1,3-Dichlorobenzene	ND	0.50	ND	0.083	
106-46-7	1,4-Dichlorobenzene	ND	0.50	ND	0.083	
95-50-1	1,2-Dichlorobenzene	ND	0.50	ND	0.083	
5989-27-5	d-Limonene	ND	0.50	ND	0.090	
96-12-8	1,2-Dibromo-3-chloropropane	ND	0.50	ND	0.052	
120-82-1	1,2,4-Trichlorobenzene	ND	0.50	ND	0.067	
91-20-3	Naphthalene	ND	0.50	ND	0.095	
87-68-3	Hexachlorobutadiene	ND	0.50	ND	0.047	

ND = Compound was analyzed for, but not detected above the laboratory reporting limit.

MRL = Method Reporting Limit - The minimum quantity of a target analyte that can be confidently determined by the referenced method.

ALS ENVIRONMENTAL

RESULTS OF ANALYSIS

Page 4 of 4

Client: South Coast Air Quality Management District

Client Sample ID: Method Blank

Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560

ALS Sample ID: P160927-MB

Tentatively Identified Compounds

Test Code: EPA TO-15

Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13

Analyst: Evelyn Alvarez

Sample Type: Canister

Test Notes:

Date Collected: NA

Date Received: NA

Date Analyzed: 9/27/16

Volume(s) Analyzed: 1.00 Liter(s)

Canister Dilution Factor: 1.00

GC/MS Retention Time	Compound Identification	Concentration µg/m ³	Data Qualifier
<hr/> No Compounds Detected <hr/>			

ALS ENVIRONMENTAL

SURROGATE SPIKE RECOVERY RESULTS

Page 1 of 1

Client: South Coast Air Quality Management District
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister(s)
 Test Notes:

Date(s) Collected: 9/23/16
 Date(s) Received: 9/27/16
 Date(s) Analyzed: 9/27/16

Client Sample ID	ALS Sample ID	1,2-Dichloroethane-d4	Toluene-d8	Bromofluorobenzene	Acceptance Limits	Data Qualifier
		Percent Recovered	Percent Recovered	Percent Recovered		
Method Blank	P160927-MB	88	104	104	70-130	
Lab Control Sample	P160927-LCS	86	103	106	70-130	
Ermanita/ 98th Ave	P1604560-001	87	104	105	70-130	

Surrogate percent recovery is verified and accepted based on the on-column result.

Reported results are shown in concentration units and as a result of the calculation, may vary slightly from the on-column percent recovery.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 1 of 3

Client: South Coast Air Quality Management District
Client Sample ID: Lab Control Sample
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P160927-LCS

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: NA
 Date Received: NA
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 0.125 Liter(s)

CAS #	Compound	Spike Amount µg/m ³	Result µg/m ³	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
115-07-1	Propene	196	163	83	49-131	
75-71-8	Dichlorodifluoromethane (CFC 12)	188	186	99	65-117	
74-87-3	Chloromethane	200	187	94	48-132	
76-14-2	1,2-Dichloro-1,1,2,2-tetrafluoroethane (CFC 114)	204	222	109	65-122	
75-01-4	Vinyl Chloride	200	208	104	65-128	
106-99-0	1,3-Butadiene	206	271	132	62-143	
74-83-9	Bromomethane	202	234	116	65-130	
75-00-3	Chloroethane	200	221	111	69-126	
64-17-5	Ethanol	998	1000	100	57-126	
75-05-8	Acetonitrile	212	168	79	51-134	
107-02-8	Acrolein	214	168	79	55-146	
67-64-1	Acetone	1,080	984	91	57-120	
75-69-4	Trichlorofluoromethane	216	196	91	59-139	
67-63-0	2-Propanol (Isopropyl Alcohol)	418	384	92	59-129	
107-13-1	Acrylonitrile	212	198	93	64-136	
75-35-4	1,1-Dichloroethene	216	234	108	72-123	
75-09-2	Methylene Chloride	222	227	102	63-117	
107-05-1	3-Chloro-1-propene (Allyl Chloride)	218	180	83	50-141	
76-13-1	Trichlorotrifluoroethane	220	235	107	68-118	
75-15-0	Carbon Disulfide	210	178	85	55-143	
156-60-5	trans-1,2-Dichloroethene	210	213	101	69-129	
75-34-3	1,1-Dichloroethane	212	202	95	66-122	
1634-04-4	Methyl tert-Butyl Ether	216	210	97	55-128	
108-05-4	Vinyl Acetate	1,040	1110	107	66-140	
78-93-3	2-Butanone (MEK)	220	227	103	62-127	

Laboratory Control Sample percent recovery is verified and accepted based on the on-column result. Reported results are shown in concentration units and as a result of the calculation, may vary slightly.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 2 of 3

Client: South Coast Air Quality Management District
Client Sample ID: Lab Control Sample
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P160927-LCS

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: NA
 Date Received: NA
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 0.125 Liter(s)

CAS #	Compound	Spike Amount µg/m ³	Result µg/m ³	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
156-59-2	cis-1,2-Dichloroethene	218	214	98	65-125	
141-78-6	Ethyl Acetate	428	436	102	64-132	
110-54-3	n-Hexane	212	178	84	58-126	
67-66-3	Chloroform	224	217	97	68-117	
109-99-9	Tetrahydrofuran (THF)	220	207	94	64-123	
107-06-2	1,2-Dichloroethane	214	203	95	63-124	
71-55-6	1,1,1-Trichloroethane	210	224	107	68-120	
71-43-2	Benzene	226	230	102	61-110	
56-23-5	Carbon Tetrachloride	230	233	101	65-137	
110-82-7	Cyclohexane	424	451	106	68-122	
78-87-5	1,2-Dichloropropane	216	214	99	67-122	
75-27-4	Bromodichloromethane	218	237	109	71-124	
79-01-6	Trichloroethene	216	238	110	71-121	
123-91-1	1,4-Dioxane	210	254	121	67-122	
80-62-6	Methyl Methacrylate	422	469	111	76-130	
142-82-5	n-Heptane	216	222	103	67-125	
10061-01-5	cis-1,3-Dichloropropene	208	227	109	73-131	
108-10-1	4-Methyl-2-pentanone	220	223	101	66-132	
10061-02-6	trans-1,3-Dichloropropene	210	234	111	76-135	
79-00-5	1,1,2-Trichloroethane	216	242	112	73-121	
108-88-3	Toluene	218	232	106	67-117	
591-78-6	2-Hexanone	220	220	100	59-128	
124-48-1	Dibromochloromethane	220	281	128	73-132	
106-93-4	1,2-Dibromoethane	218	267	122	73-128	
123-86-4	n-Butyl Acetate	226	243	108	61-136	

Laboratory Control Sample percent recovery is verified and accepted based on the on-column result. Reported results are shown in concentration units and as a result of the calculation, may vary slightly.

ALS ENVIRONMENTAL

LABORATORY CONTROL SAMPLE SUMMARY

Page 3 of 3

Client: South Coast Air Quality Management District
Client Sample ID: Lab Control Sample
Client Project ID: Exxon Mobile Torrance Refinery ER

ALS Project ID: P1604560
 ALS Sample ID: P160927-LCS

Test Code: EPA TO-15
 Instrument ID: Tekmar AUTOCAN/Agilent 5975Binert/6890N/MS13
 Analyst: Evelyn Alvarez
 Sample Type: Canister
 Test Notes:

Date Collected: NA
 Date Received: NA
 Date Analyzed: 9/27/16
 Volume(s) Analyzed: 0.125 Liter(s)

CAS #	Compound	Spike Amount µg/m ³	Result µg/m ³	% Recovery	ALS	Data Qualifier
					Acceptance Limits	
111-65-9	n-Octane	210	213	101	67-124	
127-18-4	Tetrachloroethene	202	240	119	65-126	
108-90-7	Chlorobenzene	220	249	113	68-120	
100-41-4	Ethylbenzene	218	246	113	69-123	
179601-23-1	m,p-Xylenes	428	493	115	67-125	
75-25-2	Bromoform	228	261	114	68-153	
100-42-5	Styrene	222	261	118	68-132	
95-47-6	o-Xylene	210	243	116	67-124	
111-84-2	n-Nonane	204	186	91	60-130	
79-34-5	1,1,2,2-Tetrachloroethane	210	241	115	72-128	
98-82-8	Cumene	208	240	115	67-124	
80-56-8	alpha-Pinene	212	245	116	67-129	
103-65-1	n-Propylbenzene	204	234	115	67-125	
622-96-8	4-Ethyltoluene	214	249	116	66-128	
108-67-8	1,3,5-Trimethylbenzene	214	247	115	65-125	
95-63-6	1,2,4-Trimethylbenzene	218	264	121	62-134	
100-44-7	Benzyl Chloride	220	253	115	74-145	
541-73-1	1,3-Dichlorobenzene	228	277	121	63-133	
106-46-7	1,4-Dichlorobenzene	208	259	125	62-129	
95-50-1	1,2-Dichlorobenzene	220	276	125	62-134	
5989-27-5	d-Limonene	210	245	117	66-137	
96-12-8	1,2-Dibromo-3-chloropropane	218	243	111	71-147	
120-82-1	1,2,4-Trichlorobenzene	230	251	109	60-145	
91-20-3	Naphthalene	218	249	114	56-158	
87-68-3	Hexachlorobutadiene	230	240	104	56-139	

Laboratory Control Sample percent recovery is verified and accepted based on the on-column result.
 Reported results are shown in concentration units and as a result of the calculation, may vary slightly.